



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204100	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/26/2014
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Chariton County Sheriff's Department, 307 S. Cherry, Keytesville, MO	TIME OF INSPECTION 2:50 pm
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 02/26/14 1450

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER 49 °C

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13001 EXP. DATE 03/07/2015

☒ SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G 6815 EXP. DATE 04/03/2014

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .101

TEST 2 .102

TEST 3 .102

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	2	(.10-.14)	1	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS  
(USE OTHER SIDE IF NECESSARY).

Meets Department of Health Standards

**INSPECTING OFFICER**

SIGNATURE

*Lane P. Quinn*

PRINT FULL NAME

Lane P. Quinn

TYPE II PERMIT NUMBER/EXPIRATION DATE

240025 / 01-28-16

TELEPHONE NUMBER

(660) 385-2132

RETURN COMPLETED REPORT TO THE:

Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER:** RepCo Marketing, Inc.

**LOT NUMBER:** 13001

**EXPIRATION DATE:** March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**LANE P QUINN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/28/2014

NUMBER 240025

EXPIRES 1/28/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

80-0771 (6-10)

LAB-4 (R6-10)

	STATE OF MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES BREATH ALCOHOL PROGRAM
	<b>INSTRUMENT OPERATOR CARD</b>
<small>The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.</small>	
Operator QUINN, LANE Permit No 240025 Date Issued 1/28/2014 Date Expires 1/28/2016	

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204100  
02/26/14

TESTING OFFICER:

QUINN/L/P

OFFICER I.D.: 937

PERMIT NUMBER: 240025

EXPIRATION DATE: 01/28/16

MISCELLANEOUS DATA:

----- SUPERVISOR MODE -----

BLANK TEST	.000	14:54
INTERNAL STANDARD	VERIFIED	14:54
EXTERNAL STANDARD	.101	14:54
BLANK TEST	.000	14:55
EXTERNAL STANDARD	.102	14:55
BLANK TEST	.000	14:56
EXTERNAL STANDARD	.102	14:56
BLANK TEST	.000	14:57

N = 3

S.D. = .1

AVG. = .1016

Operator Signature

*TPR L.P. Quinn*

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204100  
02/26/14

ARREST TIME: 14:00

SUBJECT NAME:

RFI/TEST

DOB: 01/01/01 SEX: M

STATE/D.L.: MO/1111111111

ARRESTING OFFICER:

QUINN/L/P

OFFICER I.D.: 937

TESTING OFFICER:

QUINN/L/P

OFFICER I.D.: 937

PERMIT NUMBER: 240025

EXPIRATION DATE: 01/28/16

MISCELLANEOUS DATA:

----- BREATH ANALYSIS -----

BLANK TEST	.000	15:05
INTERNAL STANDARD	VERIFIED	15:06
RADIO INTERFERENCE		

Operator Signature

*TPR L.P. Quinn*

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL  
BAC DATAMASTER SERIAL NUMBER 204100  
02/26/14  
14:58

## --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

## PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~  
~{|}abcdefghijklmnopqrstuvwxyz{|}~

Operator Signature

*TPR L.P. Quinn*